

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130115

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATA MASTER cdm 130115

AUGUST 27, 2009

OPERATOR'S NAME:

SUBJECT'S FIRST NAME/MI :
NO

O.I.L. #: 012345
DEPT/AGENCY: NKN
REPORTS: 012345
TEST TYPE: U
R.D. TARGET VALUE: .076 ✓
ALCO S/N: X124623

BREATH ANALYSIS

.076 ADJUSTED FOR 29.49 in ✓
ALCO TARGET .074 21:04
BLANK TEST .000 21:05
INTERNAL STANDARD VERIFIED 21:05
ALCO IN 29.51 in .075 21:05
BLANK TEST .000 21:06
SUBJECT SAMPLE .075 21:07
BLANK TEST .000 21:08
ALCO IN 29.51 in .075 21:08
BLANK TEST .000 21:09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATA MASTER cdm 130115

AUGUST 27, 2009 ✓
TIME 21:09

PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 43c
BREATH TUBE: 40c
BAROMETER: 29.51 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

ALCO IN 29.51 in .075 21:08
ALCO IN 29.51 in .075 21:08
ALCO IN 29.51 in .075 21:08

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

AUG 29 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130115 ✓
Name	DOUG E. HARRIS	ID#	5530 Date AUGUST 27, 09
Agency	BRISTOL BAY BOROUGH POLICE-NKN1	Phone #	907-246-4222
Instrument Location	BBBPD KING SALMON, AK		✓
Alco S/N		Target Value	High Pressure
X124623 ✓		.076 ✓	
Alco Test Values		1 st Alco	2 nd Alco
		.075 ✓	.075 ✓
		8/31/09	✓
Signature		Doug E. Harris ✓	

(Do Not write in this space)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of Sept, 2009.

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)

